



## Proficiency testing for veterinary laboratories

Distribution Date: 02/07/2013

### Results tabulation for 11140/BA PT0087: Salmonella in animal feed

Culture	13/1623	13/1624	13/1625	13/1626	13/1627
	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type
Intended	Salmonella Dublin (approx 300 cfu per vial)	Salmonella Mbandaka (approx 150 cfu per vial)	Salmonella Typhimurium + E. coli + Pseudomonas aeruginosa	Salmonella Typhimurium + E. coli + Pseudomonas aeruginosa	Salmonella Senftenberg (approx 700 cfu per vial)
	Group D	Group C	Group B	Group B	Group E
1002	S. Havana;	S. Braenderup	S. Typhimurium	S. Typhimurium	S. Senftenberg
1033	Group D	Group C	Group B	Group B	Group E
1127	S. enteritidis	S. paratyphi	S. typhimurium	S. typhimurium	Group E
1133	Nil Return				
1143	Group D	Negative	Group B	Group B	Group E
1320	Salmonella sp	Salmonella sp	Salmonella sp	Salmonella sp	Salmonella sp
1349	S. Enteritidis	S. Mbandaka	S. Typhimurium	S. Typhimurium	S. Senftenberg
1695	Present	Present	Present	Present	Present
2015	Group D (09)	Group C (07)	Group B (04/05)	Group B (04/05)	Group E (019)
2132	S. Dublin	S. Mbandaka	S. Typhimurium	S. Typhimurium	S. Senftenberg
2664	Detected	Detected	Detected	Detected	Detected

Intended results were determined using Rappaport Broth.

Sample 13/1625 = 13/1626

#### Comments provided by Gordon Bennett, AHVLA Lasswade

Generally a good set of isolations with all laboratories submitting a correct version of the Intended Results for sample 13/1625, 13/1626 and 13/1627.

Labs 1002 and 1127 reported the incorrect serotype or serovar for samples 13/1623 and 13/1624.

Lab 1143 failed to isolate salmonella in sample 13/1624.

Lab 1349 correctly grouped all isolates but incorrectly serotyped sample 13/1623.



0004

Version: 21/12/2012

Paula Johnson  
Scheme Coordinator  
Animal Health and Veterinary Laboratories Agency  
Quality Assurance Unit  
The Elms, College Road  
Sutton Bonington  
Loughborough  
Leicestershire  
LE12 5RB  
UK  
Authorised: 13/08/2013