



Proficiency testing for veterinary laboratories

Results tabulation for 10898/BA PT0087: Salmonella in animal feed

Distribution Date: 08/01/2013

Culture	13/1110	13/1111	13/1112	13/1113	13/1114
	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type
Intended	E coli + Pseudomonas aeruginosa	S Tennessee (approx 2 x 10 ²)	S Tennessee (approx 2 x 10 ²)	S Kedougou (approx 2 x 10 ²)	E coli + Pseudomonas aeruginosa
	Negative	Group C	Group C	Group G	Negative
1002	Not detected	S. Tennessee	S. Tennessee	S. Kedougou	Not detected
1033	-	Group C1	Group C1	Group G	-
1127	-ve	+ve group C	+ve group C	+ve group G	-ve
1133	Not Salmonella	Group C	Group C	Group G	Not Salmonella
1143	Negative	Group C	Group C	Group G	Negative
1349	Nil Return				
1388	Not detected	Detected	Detected	Detected	Not detected
1695	Absent	Present	Present	Present	Absent
1854	Negative	S. Tennessee	S. Tennessee	S. Kedougou	Negative
1901	Negative	C1 group	C1 group	G-S.Kedougou	Negative
1913	negative	positive, C1	positive, C1	positive	negative
2103	Not detected	Group C1 (6,7)	Group C1 (6,7)	present	Not detected
2132	No Salmonella	S.tennessee	S.tennessee	S.kedougou	No Salmonella
2167	Nil Return				
2664	Detected	Detected	Detected	Detected	Not Detected

Intended results were determined using Rappaport Broth.

Sample 13/1110 = 13/1114 and Sample 13/1111 = 13/1112

Comments provided by Alastair Smith, AHVLA Preston

Samples were sent to 15 laboratories, results were returned by 13 laboratories.

Twelve laboratories reported the correct positive/negative results for all five samples.

Laboratory 2664 reported an incorrect positive result for sample 13/1110.

Laboratories 1349 and 2167 each recorded a nil return for this distribution.



Version: 02/12/2011

Paula Johnson
Animal Health and Veterinary Laboratories Agency
Quality Assurance Unit
The Elms, College Road
Sutton Bonington
Loughborough
Leicestershire
LE12 5RB
UK
Authorised: 07/02/2013