



Proficiency testing for veterinary laboratories

Distribution Date: 02/10/2018

Results tabulation for 13519/BA PT0087: Salmonella in animal feed

Culture	18/1863	18/1864	18/1865	18/1866	18/1867
	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type	Salmonella Group/Type
Intended	Salmonella Infantis (approx. 5000 cfu per vial)	Salmonella Enteritidis (approx. 3000 cfu per vial) + Pseudomonas aeruginosa	Citrobacter youngae + Pseudomonas aeruginosa	Salmonella Agona (approx. 4000 cfu per vial)	Salmonella Enteritidis (approx. 7000 cfu per vial)
	Group C	Group D	Negative	Group B	Group D
1002	C1	D1	Not detected	B	D1
1127	Group C1	Group D	Negative	Group B	Group D
1133	Group C	Group D	Negative	Group B	Group D
1143	Group C	Group D	Negative	Group B	Group D
1388	Detected	Detected	Not detected	Detected	Detected
1487	C	D/Enteritidis	Negative	B	D/Enteritidis
2132	S. Infantis	S. Enteritidis	Not Salmonella	S. Agona	S. Enteritidis
2167	S. Infantis	S. Enteritidis	-ve	S. Derby	S. Enteritidis
2396	Detected O6/O7	Detected O9	Not detected	Detected O4	Detected O9
2664	DETECTED	DETECTED	Not Detected	DETECTED	DETECTED
2897	Detected Gp C	Detected Gp D	Not Detected	Detected Gp B	Detected Gp D
3010	Detected	Detected	Not Detected	Detected	Detected

Intended results were determined using Rappaport Broth.

Comments provided by Paula Johnson, Quality Assurance Unit Sutton Bonington

Results for all samples were as expected with the exception of the serotype reported by Lab 2167 for sample 18/1866, where S. Derby (Group B) was reported.



Version: 25/10/2017

Dr Anne Bollard
Deputy Scheme Coordinator
Animal and Plant Health Agency
Quality Assurance Unit
The Elms, College Road
Sutton Bonington
Loughborough
Leicestershire
LE12 5RB
UK
Authorised: 12/11/2018